



Erratum

Erratum to Evidence for the Detection of Non-Endotoxin Pyrogens by the Whole Blood Monocyte Activation Test

Nina Hasiwa^{1,2}, Mardas Daneshian¹, Peter Bruegger⁴, Stefan Fennrich⁵, Astrid Hochadel², Sebastian Hoffmann⁶, Felix E. Rivera-Mariani³, Christoph Rockel⁷, Stefanie Schindler⁸, Ingo Spreitzer⁹, Sandra Stoppelkamp⁵, Kranthi Vysyaraju³, and Thomas Hartung^{1,3}

¹CAAT-Europe, University of Konstanz, Konstanz, Germany; ²AtaX-Advice, Konstanz, Germany; ³Johns Hopkins University, Bloomberg School of Public Health, CAAT, Baltimore, USA; ⁴Novartis, Basel, Switzerland; ⁵Department of Thoracic, Cardiac and Vascular Surgery, Tübingen University Hospital, Tübingen, Germany; ⁶seh consulting + services, Paderborn, Germany; ⁷Enzler Hygiene AG, Center for Hygiene, Pratteln, Switzerland; ⁸Animalfree Research, Zürich, Switzerland; ⁹Paul-Ehrlich Institut (PEI), Langen, Germany

In this t⁴ report which appeared in ALTEX (2013), 30(2), 169-208, PMID: 23665806 there were errors in Table 14.

Tab. 14 (revised)

Tab. 14: Result of LAL-testing of the 10 spiked drugs used in the MAT validation study

| NIPH ^a | | “Truth” | | Σ |
|--------------------|---|---------|----|----|
| | | - | + | |
| PM | - | 12 | 4 | 16 |
| | + | 0 | 8 | 8 |
| Σ | | 12 | 12 | 24 |
| NIBSC ^b | | “Truth” | | Σ |
| | | - | + | |
| PM | - | 4 | 2 | 6 |
| | + | 8 | 10 | 18 |
| Σ | | 12 | 12 | 24 |

^asensitivity: 67%; specificity: 100%

^bsensitivity: 83%; specificity: 33%